

National Medical Supplies Fund
General Procurement and Contracting Department
Specification of Haemodialysis Kit for Pediatric

Appendix (3)

#	ISSUE	Specification of Haemodialysis Kits for Pediatric
1	Dialyzer	Membrane Polysulphone or International Known Material Sterilization UV, Or steam sterilization. Approved by International Quality Certificate Efficiency Highly Efficient Dialyser.
2	Blood Line	Chambers Diameter of Chamber 6.4 mm Inner diameter pump segment 6.4 Sterilization Sterile and Non pyogenic Material Flexible Remarks It must have spike and emergency nop Universal i.e can be use in all machine
3	A.V Fistula	<u>(A) Arterial</u> Diameter 17 mm or equivalent. Sterilization Sterile and Non pyogenic Tube length 150mm or equivalent Remarks It must be sharp
4	A.V Fistula	<u>(B) Venous</u> Diameter 17 mm or equivalent. Sterilization Sterile and Non pyogenic Tube length 150mm or equivalent Remarks It must be sharp
5	Bicarbonate powder	Bicarbonate - cartridge (soft edge of the cartridge) 600g - 750g weight powder Bicarbonate - bags 600g - 750g weight powder
Remark: ** The Haemodialysis kit price should include : 1. The Dialyzer. 2. Arteria & Venous Blood Line with Arterial & Venous Fistula. 3. Bicarbonate Powder.		