

Specification of Haemodialysis Kits			
NO	ISSUE		
1	Dialyzer	Membrane	polysulphone
		Sterilization	By Gamma - irradiation Or any accepted Sterilization Method
		Approved by	European or American FDA M Or International Approved Quality
2	Blood Line	Chambers	Diameter of Chamber 22 mm
		Sterilization	Sterile and Non pyogenic
		Material	Flexible
		Remarks	It must have spike and emergency nop
			Universal i.e can be use in all machine
3	A.V Fistula	(A) Arterial	
		Diameter	16 mm for adult
		Sterilization	Sterile and Non pyogenic
		Tube length	150mm or equivalent
		Remarks	It must be sharp
4	A.V Fistula	(B) Venous	
		Diameter	15 - 16 mm for adult
		Sterilization	Sterile and Non pyogenic
		Tube length	150mm or equivalent
		Remarks	It must be sharp
5	Bicarbonate powder	387,200	Bicarbonate - cartridge (soft edge of the cartridge) 600g - 750g weight powder
		96,800	Bicarbonate - bags 600g - 750g weight powder
Remark:- ** The haemodialysis kit price must included : 1/ The Dialyzer Surface area. 2/ Blood Line. 3/ A.V Fistula Arterial . 4/ A.V Fistula Venous. 5/ Bicarbonate powder.			